

# Agilent 34138A Test Lead Set

## Operating Instructions

This Test Lead Set is designed for use with the Agilent 34xxxA Multimeters, and also the U2741A, U3401A, and U3402A Multimeters. Do not use with the Agilent 3458A.

### Assistance

For technical assistance, contact your nearest Agilent Sales Office. Or visit the Agilent web site at **[www.agilent.com/find/assist](http://www.agilent.com/find/assist)** for further information.

Agilent Technologies, Inc.  
900 S. Taft Avenue  
Loveland, CO 80537 U.S.A.

### Symbols



Refer to instrument manual for additional safety information.

CAT II IEC Measurement Category II. Refer to instrument manual for additional information.

### Ratings

### Quantity

Test Leads – 1000V, 15A	2 pieces
Fine Tip Probe Attachments – 300V, 3A	2 pieces
Mini Grabber Attachment – 300V, 3A	1 piece
SMT Grabber Attachments – 300V, 3A	2 pieces

### Operation

- Refer to the multimeter user's guide for further information.
- The Fine Tip, Mini Grabber, and SMT Grabber attachments plug onto the probe end of the Test Leads.

**WARNING:** If Test Lead Set is used in a manner not specified by Agilent Technologies, the protection provided by the Test Lead Set may be impaired. *Instrument damage or personal injury may result.*

### Maintenance

If any portion of the Test Lead Set is worn or damaged, do not use; replace with a new Test Lead Set.



**Agilent Technologies**



E0213

34138-90000