


# Technical Data Sheet

# Rosenberger

RPC-1.35

Open Circuit  
Jack

P9K12L-000D3



All dimensions are in mm; tolerances according to ISO 2768 m-H

## Interface

According to

IEC 61169-65

## Documents

Application note

AN001 "Calibration Services"

## Material and plating

### Connector parts

Center conductor  
Outer conductor  
Dielectric

### Material

CuBe  
CuBe or equiv.  
PMP

### Plating

Gold, min. 1.27µm  
Gold, min. 1.27µm

## Electrical data

Frequency range DC to 90 GHz

Error from nominal phase<sup>1</sup>

≤ 3.0°, DC to 40 GHz
≤ 4.0°, 40 GHz to 65 GHz
≤ 5.0°, 65 GHz to 90 GHz

<sup>1</sup> The nominal phase is defined by the Offset Delay, the Offset Loss and the Fringing Capacitances.

## Mechanical data

Mating cycles	≥ 3000
Maximum torque	1.65 Nm
Recommended torque	0.90 Nm
Gauge	0.003 mm to 0.020 mm

## General standard definitions

For proper operation the vector network analyzer (VNA) needs a model describing the electrical behaviour of this calibration standard. The different models, units, and terms used will depend on the VNA type and they will have to be entered into the VNA. All values are based on typical geometry and plating.

Offset $Z_o$ / Impedance / $Z_o$	50 $\Omega$
Offset Delay	16.678 ps
Length (electrical) / Offset Length	5.00 mm
Offset Loss	5.95 G $\Omega$ /s
Loss	0.0172 dB/ $\sqrt{\text{GHz}}$
Fringing Capacitances <sup>2</sup>	

<sup>2</sup> Fringing Capacitances are determined individually for each open circuit and are documented in a Calibration Certificate.

## Environmental data

Operating temperature range <sup>3</sup>	+ 20 °C to +26 °C
Rated temperature range of use <sup>4</sup>	0 °C to +50 °C
Storage temperature range	- 40 °C to +85 °C

RoHS	compliant
------	-----------

<sup>3</sup> Temperature range over which these specification are valid.

<sup>4</sup> This range is underneath and above the operating temperature range, within the open circuit is fully functional and could be used without damage.

## Declaration of calibration options

Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co. KG  
RF\_35/09.14/6.2

Technical Data Sheet				Rosenberger									
RPC-1.35		Open Circuit Jack		P9K12L-000D3									
<p><b>Factory Calibration</b></p> <p>Standard delivery for this calibration standard includes a Factory Calibration. The Calibration Certificate issued reports individual calibration results, <b>traceable to Rosenberger standards</b>, national / international standards are not available. Model based standard definitions are individually optimized and reported in an Agilent/Keysight, Rohde &amp; Schwarz and Anritsu compatible VNA format.</p> <p><b>Accredited Calibration</b></p> <p>Not available.</p> <p><i>For further, more detailed information see application note AN001 on the Rosenberger homepage.</i></p> <p><b>Calibration interval</b></p> <table><tr><td>Recommendation</td><td>12 months</td></tr></table> <p><b>Packing</b></p> <table><tr><td>Standard</td><td>1 pce in box</td></tr><tr><td>Weight</td><td>6.5 g/pce</td></tr></table>								Recommendation	12 months	Standard	1 pce in box	Weight	6.5 g/pce
Recommendation	12 months												
Standard	1 pce in box												
Weight	6.5 g/pce												
<p>While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.</p>													
Draft		Date		Rev.		Engineering change number							
Approved		Date		Name		Date							
Marcel Panicke		16.11.18		Lars Ramtke		15.11.19							
b00		19-2148		Marion Striegler		15.11.19							
Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany <a href="http://www.rosenberger.de">www.rosenberger.de</a>				Tel. : +49 8684 18-0 Email : <a href="mailto:info@rosenberger.de">info@rosenberger.de</a>			Page 3 / 3						