


Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co. KG
RF_35/09.14/6.2

Technical Data Sheet		Rosenberger
RPC-3.50 / RPC-2.92 RPC-2.40 / RPC-1.85	Torque Wrench	03W021-000
 Labeled with Rosenberger-Logo, ident.no. and torque		
All dimensions are in mm; tolerances according to ISO 2768 m-H		
Documents Application noteAN001 "Calibration Services"		
Material and plating Tooling parts Spanner MaterialStainless steel PlatingNickel, 2.5-5 µm		
Mechanical data Torque setting Width across flats 0.90 Nm ± 0.05 Nm (force applied in the middle of handle) 8 mm		
Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de		Tel. : +49 8684 18-0 Email : info@rosenberger.de
		Page 1 / 2

Dieses Dokument ist urheberrechtlich geschützt • This document is protected by copyright • Rosenberger Hochfrequenztechnik GmbH & Co. KG
RF_35/09.14/6.2

Technical Data Sheet				Rosenberger			
RPC-3.50 / RPC-2.92 RPC-2.40 / RPC-1.85		Torque Wrench		03W021-000			
Declaration of calibration options Factory Calibration Standard delivery for this torque wrench includes a Factory Calibration. The Calibration Certificate issued reports individual mechanical calibration results, traceable to national / international standards. Accredited Calibration Not available. For further, more detailed information see application note AN001 on the Rosenberger homepage. Calibration interval Recommendation12 months Packing Standard1 pce in box Weight121.5 g/pce							
While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.							
Draft	Date	Approved	Date	Rev.	Engineering change number	Name	Date
Herbert Babinger	15.10.14	Martin Moder	15.10.14	e00	14-1492	Herbert Babinger	15.10.14
Rosenberger Hochfrequenztechnik GmbH & Co. KG P.O.Box 1260 D-84526 Tittmoning Germany www.rosenberger.de					Tel. : +49 8684 18-0 Email : info@rosenberger.de		Page 2 / 2